

Hu-CD45RA/CD62L/CD3/CD8

4 Color FCM Reagent: sc-2909



BACKGROUND

Human CD45RA/CD62L/CD3/CD8: sc-2909 is a direct immunofluorescence reagent formatted to identify and determine the percentage of mature T cells and suppressor/cytotoxic (naive and memory) T lymphocyte subsets in erythrocyte-lysed whole blood, based on cell-surface antigen expression. CD45 is a major leukocyte cell surface molecule that is essential for the activation of T and B lymphocytes (1,2). In T cells, the alternative splicing of CD45 is regulated so that naive or unprimed T cells predominantly express CD45RO positive isoforms and switch to expression of CD45RA upon activation (3). CD62L is present on a subset of normal peripheral blood B lymphocytes and on most circulating T cells (4). CD3 identifies T lymphocytes and non-covalently associates with either α/β or γ/δ TCR, which recognizes antigens associated with the MHC (5). CD8 identifies suppressor/cytotoxic T lymphocytes and binds class I MHC molecules, which enhances the activation of resting T lymphocytes (6).

Antigen Expression	Cell Type Identified
CD3+	Mature T Cells
CD3+ CD8+	Suppressor/Cytotoxic T Cells
CD3+ CD8+ CD45RA+ CD62L+	Suppressor/Cytotoxic Naive T Cells
CD3+ CD8+ CD45RA+ CD62L-	Suppressor/Cytotoxic Memory T Cells
CD3+ CD8+ CD45RA- CD62L-	Suppressor/Cytotoxic Memory T Cells

STORAGE

Store at 4° C. Do not freeze. Stable for one year from the date of shipment. Protect reagents from prolonged exposure to light.

PRODUCT

Supplied in 1.0 ml of PBS containing 0.1% azide and 0.1% gelatin. Sufficient for 50 tests. This product has been titrated for optimal performance. Recommended use is 20 μ L per test (1×10^6 cells). **For research use only. Not for use in diagnostic procedures.**

INSTRUMENT

Human CD45RA/CD62L/CD3/CD8: sc-2909 is recommended for use with a dual laser Flow Cytometer fitted with appropriate acquisition and analysis software, such as the FACSCalibur™ Flow Cytometer fitted with CellQuest™ Software by Becton Dickinson.

The flow cytometer must be equipped with 635 nm and 488 nm lasers and must be capable of detecting light scatter (forward and side) and four-color fluorescence with emission detectable in four ranges: 515-545 nm, 562-607 nm, >650 nm and 652-668 nm, and it must be able to threshold and discriminate using the >650 channel.

Antigen	Clone	Isotype	Label*	Detection Range (nm)
CD45RA	4KB5	IgG ₁	FITC	515-545
CD62L	DREG56	IgG ₁	PE	562-607
CD3	UCH-T1	IgG ₁	PE-Cy5	>650
CD8	HIT8a	IgG ₁	APC	652-668

*Fluorescent labels include FITC: Fluorescein isothiocyanate; PE: phycoerythrin; PE-Cy5: phycoerythrin-cyanin 5; APC: allophycocyanin

ISOTYPE CONTROL

sc-2909 CON (IgG₁ FITC/IgG₁ PE/IgG₁ PE-Cy5/IgG₁ APC) is the isotype matched negative control for this system and is suitable for 50 tests.

REFERENCES

- Charbonneau, H., Tonks, N.K., Walsh, K.A., and Fischer, E.H. 1988. The leukocyte common antigen (CD45): a putative receptor-linked protein tyrosine phosphatase. *Proc. Nat. Acad. Sci. USA* **85**: 7182-7186.
- Kishihara, K., Penninger, J., Wallace, V.A., Kundig, T.M., Kawai, K., Wakeham, A., Timms, E., Pfeffer, K., Ohashi, P.S., and Thomas, M.L. 1993. Normal lymphocyte development but impaired cell maturation in CD45-exon6 protein tyrosine phosphatase-deficient mice. *Cell* **74**: 143-156.
- Akbar, A.N., Terry, L., Timms, A., Beverley, P.C., and Janossy, G. 1988. Loss of CD45R and gain of UCHL1 reactivity is a feature of primed T cells. *J. Immunol.* **140**: 2171-2178.
- Kansas, G.S., Wood, G.S., Fishwild, D.M., and Engleman, E.G. 1985. Functional characterization of human T lymphocyte subsets distinguished by monoclonal anti-leu-8. *J. Immunol.* **134**: 2995-3002.
- Exley, M., Terhorst, C., and Wileman, T. 1991. Structure, assembly and intracellular transport of the T cell receptor for antigen. *Semin. Immunol.* **3**: 283-297.
- Gallagher, P.F., Fazekas de St Groth, B., and Miller, J.F. 1989. CD4 and CD8 molecules can physically associate with the same T-cell receptor. *Proc. Natl. Acad. Sci. USA* **86**: 10044-10048.