Hu-CD3/CD8/CD45 3 Color FCM Reagent: sc-2936



BACKGROUND

Human CD3/CD8/CD45: sc-2936 is a direct immuno-fluorescence reagent formatted to identify and determine the percentage of mature human T lymphocytes and suppressor/cytotoxic T lymphocytes in erythrocyte-lysed whole blood, based on cell-surface antigen expression. CD3 identifies T lymphocytes and non-covalently associates with either α/β or γ/δ TCR (1). CD45 is a major leukocyte cell surface molecule and a member of the leukocyte common antigen (LCA) family (2). CD8 identifies suppressor/cytotoxic T lymphocytes and binds class I MHC molecules, which enhances the activation of resting T lymphocytes (4). CD3+CD8+ and CD3+CD4+ percentages are used to characterize and monitor some forms of immunodeficiency and autoimmune disease (5,6).

Antigen Expression	Cell Type Identified	
CD3+	Mature T Cells	
CD3+ CD8+	Suppressor/Cytotoxic T Cells	

STORAGE

Store at 4° C. Do not freeze. Stable for one year from the date of shipment. Protect reagents from prolonged exposure to light.

PRODUCT

Supplied in 1.0 ml of PBS containing 0.1% azide and 0.1% gelatin. Sufficient for 50 tests. This product has been titrated for optimal performance. Recommended use is 20 uL per test (1x106 cells). For research use only. Not for use in diagnostic procedures.

INSTRUMENT

Human CD3/CD8/CD45: sc-2936 is recommended for use with either a single or dual laser Flow Cytometer fitted with appropriate acquisition and analysis software, such as the FACSCalibur™ Flow Cytometer fitted with CellQuest™ Software by Becton Dickinson.

The flow cytometer must be equipped with a 488 nm laser and must be capable of detecting light scatter (forward and side) and three-color fluorescence with emission detectable in three ranges: 515-545 nm, 562-607 nm and >650 nm, and it must be able to threshold and discriminate using the >650 channel.

Antigen	Clone	Isotype	Label*	Detection Range (nm)
CD3	UCH-T1	IgG ₁	FITC	515-545
CD8	НІТ8а	IgG ₁	PE	562-607
CD45	2D-1	IgG ₁	PE-Cy5	>650

^{*}Fluorescent labels include FITC: Fluorescein isothiocyanate; PE: phycoerythrin; PE-Cy5: phycoerythrin-cyanin 5.

ISOTYPE CONTROL

sc-2936 CON (IgG₁ FITC/IgG₁ PE/IgG₁ PE-Cy5) is the isotype matched negative control for this system and is suitable for 50 tests.

REFERENCES

- 1. Exley, M., Terhorst, C., and Wileman, T. 1991. Structure, assembly and intracellular transport of the T cell receptor for antigen. Semin. Immunol. 3: 283-297.
- 2. Charbonneau, H., Tonks, N.K., Walsh, K.A., and Fischer, E.H. 1988. The leukocyte common antigen (CD45): a putative receptor-linked protein tyrosine phosphatase. Proc. Nat. Acad. Sci. USA 85: 7182-7186.
- 3. Reichert, T., DeBruyere, M., Deneys, V., Totterman, T., Lydyard, P., Yuksel, F., Chapel, H., Jewell, D., Van Hove, L., Linden, J., *et al.* 1991. Lymphocyte subset reference ranges in adult Caucasians. Clin. Immunol. Immunopathol. 60: 190-208.
- 4. Gallagher, P.F., Fazekas de St Groth, B., and Miller, J.F. 1989. CD4 and CD8 molecues can physically associate with the same T-cell receptor. Proc. Natl. Acad. Sci. USA 86: 10044-10048.
- 5. Foucar, K. and Goeken, J.A. 1982. Clinical Applications of immunologic techniques to the diagnosis of lymphoproliferative and immunodeficiency disorders. Lab. Med. <u>13</u>: 403-413.
- 6. Smolen, J.S., Chused, T.M., Leiserson, W.M., Reeves, J.P., Alling, D., and Steinberg, A.D. 1982. Heterogeneity of immunoregulatory T-cell subsets in systemic lupus erythematosus. Correlation with clinical features. Am. J. Med. 2: 783-790.