

# Hu-CD5/CD19

## 2 Color FCM Reagent: sc-3956



### BACKGROUND

Human CD5/CD19: sc-3956 is a direct immunofluorescence reagent formatted to identify and determine the percentage of human T lymphocytes and B lymphocytes in erythrocyte-lysed whole blood, based on cell-surface antigen expression. CD5 is expressed on virtually all T lymphocytes in thymus and peripheral blood and on a subset of B lymphocytes (1). CD5 may be involved in T cell activation, when present (2). CD19 is present on human B lymphocytes during all stages of B cell maturation, but is lost on plasma cells (3). The total population of T lymphocytes and B lymphocytes are used to characterize and monitor some forms of immunodeficiency and autoimmune disease (4,5).

Antigen Expression	Cell Type Identified
CD5+	T Lymphocytes
CD19+	B Lymphocytes

### STORAGE

Store at 4° C. Do not freeze. Stable for one year from the date of shipment. Protect reagents from prolonged exposure to light.

### PRODUCT

Supplied in 1.0 ml of PBS containing 0.1% azide and 0.1% gelatin. Sufficient for 50 tests. This product has been titrated for optimal performance. Recommended use is 20 uL per test (1x10<sup>6</sup> cells). **For research use only. Not for use in diagnostic procedures.**

### INSTRUMENT

Human CD5/CD19: sc-3956 is recommended for use with either a single or dual laser Flow Cytometer fitted with appropriate acquisition and analysis software, such as the FACSCalibur™ Flow Cytometer fitted with CellQuest™ Software by Becton Dickinson.

The flow cytometer must be equipped with a 488 nm laser and must be capable of detecting light scatter (forward and side) and two-color fluorescence with emission detectable in two ranges: 515-545 nm and 562-607 nm.

Antigen	Clone	Isotype	Label*	Detection Range (nm)
CD5	UCH-T2	IgG <sub>1</sub>	FITC	515-545
CD19	SJ25C1	IgG <sub>1</sub>	PE	562-607

\*Fluorescent labels include FITC: Fluorescein isothiocyanate; PE: phycoerythrin

### ISOTYPE CONTROL

sc-3956 CON (IgG<sub>1</sub> FITC/IgG<sub>1</sub> PE) is the isotype matched negative control for this system and is suitable for 50 tests.

### REFERENCES

- Knowles, R.W. Immunohistochemical analysis of the T-cell specific antigens. In: Reinherz, E.L., Haynes, B.F., Nadler, L.M., Bernstein, I.D. eds. *Leukocyte Typing II: Human T Lymphocytes*. New York, NY: Springer-Verlag; 1986: 303-313.
- Bierer, B.E., Nishimura, Y., Burakoff, S.J., and Smith, B.R. 1988. Phenotypic and functional characterization of human cytolytic T cells lacking expression of CD5. *J. Clin. Invest.* **81**: 1390-1397.
- Dörken, B., Möller, P., Pezzutto, A., Schwartz-Albiez, R., and Moldenhauer, G. B-cell antigens: CD19. In: Knapp, W., Dörken, B., Gilks, W.R. *et al.*, eds. *Leukocyte Typing IV: White Cell Differentiation Antigens*. New York, NY: Oxford University Press; 1989: 34-36.
- Foucar, K. and Goeken, J.A. 1982. Clinical Applications of immunologic techniques to the diagnosis of lymphoproliferative and immunodeficiency disorders. *Lab. Med.* **13**: 403-413.
- Smolen, J.S., Chused, T.M., Leiserson, W.M., Reeves, J.P., Alling, D., and Steinberg, A.D. 1982. Heterogeneity of immunoregulatory T-cell subsets in systemic lupus erythematosus. Correlation with clinical features. *Am. J. Med.* **2**: 783-790.