Mo-CD4/CD8 2 Color FCM Reagent: *sc-3979*



BACKGROUND

Mouse CD4/CD8: sc-3979 is a direct immunofluorescence reagent formatted to identify and determine the percentage of mature helper/inducer and suppressor/cytotoxic T cells in erythrocytelysed whole blood, based on cell-surface antigen expression. In immune deficiency states, helper T cells decline and suppressor T cells increase. CD4 identifies helper/inducer T lymphocytes and binds class II MHC molecules (1). CD4 is also the primary receptor for HIV (2). As HIV progresses, infected individuals exhibit a steady decrease in helper/inducer lymphocytes 3,4). CD8 identifies suppressor/cytotoxic T lymphocytes (5,6) and binds class I MHC molecules, resulting in enhanced activation of resting T lymphocytes (1). Beginning at seroconversion and during the first stages of HIV infection, the number of CD8+ T lymphocytes increases, while the number of CD4+ T lymphocytes decreases (7). At the onset of AIDS, CD3+, CD4+ and CD8+ T lymphocyte levels decline, but in the late stages of AIDS, the remaining lymphocytes are CD8+ (7).

Antigen Expression	Cell Type Identified
CD4+	Helper/Inducer T Cells
CD8+	Suppressor/Cytotoxic T Cells

STORAGE

Store at 4° C. Do not freeze. Stable for one year from the date of shipment. Protect reagents from prolonged exposure to light.

PRODUCT

Supplied in 1.0 ml of PBS containing 0.1% azide and 0.1% gelatin. Sufficient for 50 tests. This product has been titrated for optimal performance. Recommended use is 20 uL per test (1x106 cells). For research use only. Not for use in diagnostic procedures.

INSTRUMENT

Mouse CD4/CD8: sc-3979 is recommended for use with either a single or dual laser Flow Cytometer fittd with appropriate acquisition and analysis software, such as the FACSCalibur™ Flow Cytometer fitted with CellQuest™ Software by Becton Dickinson.

The flow cytometer must be equipped with a 488 nm laser and must be capable of detecting light scatter (forward and side) and two-color fluorescence with emission detectable in two ranges: 515-545 nm, 562-607 nm.

Antigen	Clone	Isotype	Label*	Detection Range (nm)
CD4	H129.19	rat IgG _{2a}	FITC	515-545
CD8	53-6.7	rat IgG _{2a}	PE	562-607

^{*}Fluorescent labels include FITC: Fluorescein isothiocyanate; PE: phycoerythrin

ISOTYPE CONTROL

sc-3979 CON (rat IgG_{2a} FITC/rat IgG_{2a} PE) is the isotype matched negative control for this system and is suitable for 50 tests.

REFERENCES

- 1. Gallagher, P.F., Fazekas de St. Groth, B., and Miller, J.F. 1989. CD4 and CD8 molecues can physically associate with the same T-cell receptor. Proc. Natl. Acad. Sci. USA 86: 10044-10048.
- 2. Dalgleish, A.G., Beverley, P.C.L., Clapham, P.R., Crawford, D.H., Greaves, M.F., and Weiss, R.A. 1984. The CD4 (T4) antigen is an essential component of the receptor for the AIDS retrovirus. Nature 312: 763-767.
- 3. Landay, A., Ohlsson-Wilhelm, B., and Giorgi, J.V. 1990. Application of flow cytometry to the study of HIV infection. AIDS 4: 479-497.
- 4. Giorgi, J. and Hultin, L. 1990. Lymphocyte subset alterations and immunophenotyping by flow cytometry in HIV disease. Clin. Immunol. Newsletts. <u>10</u>: 55-62.
- 5. Ledbetter, J.A., Evans, R.L., Lipinski, M., Cunningham-Rundles, C., Good, R.A., and Herzenberg, L.A. 1981. Evolutionary conservation of surface molecules that distinguish T-lymphocyte helper/inducer and T cytotoxic/suppressor subpopulations in mouse and man. J. Exp. Med. <u>153</u>: 310-323.
- 6. Evans, R.L., Wall, D.W., Platsoucas, C.D., *et al.* 1981. Thymus-dependent membrane antigens in man: Inhibition of cell-mediated lympholysis by monoclonal antibodies to the TH2 antigen. Proc. Natl. Acad. Sci. USA <u>78</u>: 544-548.
- 7. Giorgi, J.V., Ho, H.N., Hirji, K., Chou, C.C., Hultin, L.E., O'Rourke, S., Park, L., Margolick, J.B., Ferbas, J., and Phair, J.P. 1994. CD8+ lymphocyte activation at human immunodeficiency virus type 1 seroconversion: development of HLA-DR+ CD38-CD8+ cells is associated with subsequent stable CD4+ cell levels. The Multicenter AIDS Cohort Study Group. J. Infect. Dis. 170: 775-781.